

ADA CONTROL LOW & HIGH

CONTROL FOR ADENOSINE DEAMINASE(ADA) ASSAY

Introduction

ADA Control Low & High are intended to be used for the quality control check during Adenosine deaminase (ADA) assay.

Control Storage, Stability & Handling

Shelf - Life

Unopened control is stable when stored at 2 - 8°C until the expiration date printed on the vial label.

Preparation of Control & Stability

Lyophilized ADA Control Low & High should be reconstituted by adding 1.0 ml of distilled or deionized water. Close the vial and dissolve the contents of the vial by swirling gently to avoid formation of foam. **Do not shake.** Allow it to stand for 20 minutes before use. Reconstituted ADA Controls are stable for **7 days at 2 - 8°C.**

Control Target & Range Concentration

The control target and range concentration is as follows:

	Low (A434)	High (A435)
ADA (U/L)	29.6(24.6– 34.6)	153.0 (127.0 – 179.0)

Indications of Deterioration

Presence of extreme turbidity or microbial growth may indicate deterioration.

Precautions & Warnings

1. For *in vitro diagnostic* use.
2. Do not pipette by mouth.
3. Possible infectious agents in materials used for formulating the control sera have been inactivated. However, all biological products should still be handled in accordance with good laboratory practices using appropriate precautions to avoid transmission of any infectious diseases.
4. Do not use the control after the expiration date printed on the box.

A.R.No.:I116

CRL-9-2015-02-001



ACCUREX BIOMEDICAL PVT. LTD.

Head Office - Mumbai. Tel.: 91(022) 67446744

Plant : G-54/55, MIDC Tarapur, Boisar - 401 506. INDIA.

Customer Support No.: 1800 209 8456

www.accurex.net / support@accurex.net

ADA CONTROL LOW & HIGH

CONTROL FOR ADENOSINE DEAMINASE(ADA) ASSAY

Introduction

ADA Control Low & High are intended to be used for the quality control check during Adenosine deaminase (ADA) assay.

Control Storage, Stability & Handling

Shelf - Life

Unopened control is stable when stored at 2 - 8°C until the expiration date printed on the vial label.

Preparation of Control & Stability

Lyophilized ADA Control Low & High should be reconstituted by adding 1.0 ml of distilled or deionized water. Close the vial and dissolve the contents of the vial by swirling gently to avoid formation of foam. **Do not shake.** Allow it to stand for 20 minutes before use. Reconstituted ADA Controls are stable for **7 days at 2 - 8°C.**

Control Target & Range Concentration

The control target and range concentration is as follows:

	Low (A434)	High (A435)
ADA (U/L)	29.6(24.6– 34.6)	153.0 (127.0 – 179.0)

Indications of Deterioration

Presence of extreme turbidity or microbial growth may indicate deterioration.

Precautions & Warnings

1. For *in vitro diagnostic* use.
2. Do not pipette by mouth.
3. Possible infectious agents in materials used for formulating the control sera have been inactivated. However, all biological products should still be handled in accordance with good laboratory practices using appropriate precautions to avoid transmission of any infectious diseases.
4. Do not use the control after the expiration date printed on the box.

A.R.No.:I116

CRL-9-2015-02-001



ACCUREX BIOMEDICAL PVT. LTD.

Head Office - Mumbai. Tel.: 91(022) 67446744

Plant : G-54/55, MIDC Tarapur, Boisar - 401 506. INDIA.

Customer Support No.: 1800 209 8456

www.accurex.net / support@accurex.net